

Certificate of Analysis

Product: Probe Wash 3
Product Number: SM-B01418351D
Shelf Life: 3 Year
Storage Conditions: 2 – 25° C
Quantity: 1 X 500 mL

Parameter	Specified Range	Units
Hypochlorite	5.5 - 6.5	%
Potassium	600 - 700	mmol/L
pH	10 - 14	pH Units

Details

Function: For use as a probe cleaning agent on the Advia Chemistry Systems.

Quality Specifications: Passed

Reagent Contents: Contains 4.5% Potassium Hydroxide, 4.0% Sodium polyacrylic acid, 4.7% Sodium Hypochlorite, and surfactant. Recommended storage is refrigerated.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Hypochlorite	OEM Standard	Thiosulfate Titration
Potassium	NIST 918b	Flame Photometry
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.