

Certificate of Analysis

Product: Cuvette Conditioner
Product Number: SM-B01417951D
Shelf Life: 2 Years
Storage Conditions: 18 – 25° C
Quantity: 1000 mL

Parameter	Specified Range	Units
Appearance	Clear to light yellow, free of visible particulate matter	
Potassium	13.0 ± 0.3	mmol/L
Conductivity	1.20 ± 0.05	mS/cm ³
pH	5.0 ± 0.5	pH units

Details

Function: For use as a cuvette conditioner agent on the Advia Chemistry System.

Quality Specifications: Passed


Reagent Contents: An aqueous solution containing <0.5% K Sorbic Acid, <0.1% Citric Acid, and surfactant.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Potassium	NIST 918b	IL 943 (Flame Photometry)
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.