

Certificate of Analysis

Product: ISE Detergent
Product Number: SM-B01417452D
Shelf Life: 3 Years
Storage Conditions: 2 – 8° C
Quantity: 2 x 100 mL

Parameter	Specified Range		Units
Hypochlorite	5.5	- 6.5	%
Potassium	0.85	- 0.95	mmol/L
Chloride	105	- 115	mmol/L
pH	10.0	- 14.0	pH units
Conductivity	110	- 140	mS/cm

Details

Function: For use on an Advia Analyzer to rinse the pipettes.

Quality Specifications: Passed


Reagent Contents: 6% Sodium Hypochlorite. Recommended storage is refrigerated.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Hypochlorite	OEM Reagent	Thiosulfate Titration
Potassium	NIST 918b	Flame Photometry
Chloride	NIST 919b	Titrimetric
pH	Secondary Standard Traceable to NIST	pH Meter

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.